

ABSTRACT

A method for monitoring a bolus delivered to a pressure-responsive valve of a fluid infusion catheter includes delivering the bolus through the infusion catheter over a period of time lasting between approximately one minute and approximately 60 minutes and observing a measured pressure versus time during the bolus delivery.